



DW-S430 NIR

SPECTROPHOTOMETER

User Manual

Preface

**Thank you for purchasing this instrument.
This manual will show you how to use
Prolab.**

**Please read this manual carefully before
using Prolab.**

Prolab is suitable for:

- **DW-S430 NIR Spectrophotometer**

Special Statement

Please read this manual carefully before installation or operation. The company will not take responsible for any trouble or damage due to improper use. The type DW-S430 is shown as S430 in the rest of the manual.

The company has the final interpretation of this manual. Modifications of the manual due to improvements of the instrument will not be announced.

The company will conduct 12 months free repair from the date of delivery if the instrument is in strict accordance with the instructions and the transport safety specification. (Vulnerability and consumable parts are not included)

Please use our original packaging when returning the instrument for service with accessories and the warranty card.

Any chapter or images of this manual are not allowed to borrow, copy and translate to other languages without permission of the company.

Notice

1. The instrument is suitable for analysis in laboratory. If the instrument is needed outside the lab, please make the field work environment meets the environmental requirements of the laboratory.
2. Please use the original package when moving the instrument.
3. Please wait 30 minutes after turning on the instrument to make it stable.
4. When the instrument is on, the temperature of the vents on top left corner is high. Please keep the air circulating and away from the vents surface.
5. When the instrument is on, the temperature of the vents on top left corner is high. Please keep the air circulating and away from the vents surface.
6. Please make sure the fans on the left side and top left corner operate normally. If the fans are not functioning, please turn off the instrument for repairs.
7. When an error occurred by wrong operation or other machine or instrument error, shut down the instrument immediately. When the software is not operating properly, Start TaskManager to end the "Prolab.exe" process, then restart the software and the instrument.
8. DO NOT loose the screws in the monochromator. Keep the environment clean.
9. Cut the power before opening the instrument. Pay attention to the high-voltage electrical components on the left rear of the instrument.
10. Cover the instrument with dust proof if the instrument is not used for a long time.

MENU

Preface.....	1
Special Statement.....	2
Notice.....	3
MENU	4
1 Instrument Overview	6
1.1 Theory	6
1.2 Appearance	7
1.3 Performance	8
1.4 Packing List	8
2 Environmental Requirements of Software.....	9
2.1 PC Requirements	9
2.1.1 Hardware Requirements.....	9
2.2 Install Prolab	9
3 Before Use.....	11
3.1 Connect to PC	11
3.2 How To Connect	11
4 Functions	12
4.1 Modes.....	12
4.2 Interface.....	13
4.2.1 Modules.....	13
4.2.2 Modules intro.....	13
4.2.3 Icons.....	15
5 Software Operation	17
5.1 Photometric	17
5.1.1 Create Measurement.....	17
5.1.2 Start a Measurement	21
5.1.3 Data Processing	21
5.2 Wavelength Scan.....	22
5.2.1 Create a Measurement.....	23
5.2.2 Start a wavelength scan	27
5.2.3 Data Processing	27
5.2.4 Model.....	28
5.3 Time Scan.....	29
5.3.1 Create a new measurement	29
5.3.2 Start a Time Scan	34
5.3.3 Data Processing	34
5.4 Quantitation	36
5.4.1 Create a measurement.....	37
5.4.2 Build Calibration Curve.....	39
5.4.3 Measuring unknown samples	41
5.4.4 Data Processing	42

5.5	General Operation	43
5.5.1	Wavelength.....	43
5.5.2	Set 100%.....	43
5.5.3	Set 0%.....	43
5.5.4	Move sample rack	43
5.5.5	Switch T/A	44
5.5.6	Locate file	45
5.5.7	Data management	45
5.5.8	Characteristic Peak	46
5.5.9	Arithmetic	46
5.5.10	Spectrum Compare	47
5.5.11	Find Peak	49
5.5.12	Smooth	49
5.5.13	Derivative	49
5.5.14	Instrument Parameter.....	50
5.5.15	Rename & Delete Files.....	52
6	Appendix.....	53
6.1	Quantitative analysis wavelength method.....	53
6.1.1	Single Wavelength.....	53
6.1.2	Double Wavelengths.....	53
6.1.3	Triple Wavelengths	54
6.2	DETAILS ON QUANTITATIVE	54
6.2.1	Linear Working Curve (1st order)	54
6.2.2	Quadratic Working Curve (2nd order).....	55
6.2.3	The correlation coefficient.....	55
6.3	Derivative Operation on Spectrum.....	56
6.4	Smoothing	57
6.4.1	Savitzky–Golay.....	57
6.4.2	Mean smoothing	57
6.4.3	Median smoothing	57

1 Instrument Overview

1.1 Theory

The principle of spectrophotometric analysis is to show the substances absorb phenomenon on different wavelengths and do qualitative and quantitative analysis of the substances.

This instrument is measuring base on relative measurement. Choose a certain substance (Distilled water, air or the sample) as reference solution, set its transmittance to 100%. The transmittance of the sample being tested is relative to the reference solution. The transmittance and concentration of the test substance is related. Within a certain range, it is consistent with Lambert - Beer law.

$$T=I/I_0$$

$$A=KCL=-\log I/I_0$$

T :Transmittance

A: Absorbance

C: Concentration

K: Absorption coefficient of the solution

L: Length of the liquid layer in the optical path

I: Intensity of light transmitted through the test sample on the photoelectric converter

I₀: Intensity of light transmitted through the reference sample on the photoelectric converter

This instrument is widely used in medicine and health, clinical testing, biochemical, petrochemical, environmental monitoring, food production and quality control departments for qualitative and quantitative analysis. It is also as a teaching demonstration and laboratory equipment, and post-secondary institutions and related courses.


The S430 Near-Infrared Spectrometer is a grating-type near-infrared spectrometer. Wavelength range is 1000nm-1800nm. This instrument is mainly used for non-destructive testing of liquid samples. The analysis process is very convenient, Just put the cuvette filled with samples into the instrument sample compartment, click on the PC software to get the sample spectral data. The spectrum data files can be opened in other analysis softwares.

S430NIR spectrophotometer can be widely used in oil, alcohol, beverages and other liquid quality of fast nondestructive analysis.

1.2 Appearance



Fig 1.2.1 Appearance



1. Sample compartment lid 2. Power switch 3. Fuse 4. Power socket 5. USB Port

Fig 1.2.2 Left view

1.3 Performance

Wavelength Range	900-2500 nm
Wavelength Accuracy	$\pm 0.2\text{nm}$
Wavelength Reproducibility	$\leq 0.05\text{nm}$
Absorbance Noise	$< 50\mu\text{A}$
Baseline flatness	$\pm 0.001\text{A}$
Resolution	8nm
Stray Light	$< 0.1\%$
Light Source	> 10000 hours
Scanning time	< 1 minute
Communicate Port	USB2.0
Size	360×460×240mm
Weight	12kg

1.4 Packing List

Main body	1 pc
Power cable	1 pc
User manual	1 copy
Certification	1 copy
Fuse (2A)	2 pcs
10mm Quartz rectangular cuvette	1 set (2 pcs)
2mm Quartz rectangular cuvette	1 set (2 pcs)
Warranty	1 copy
Flash disk	1 pc
Software copy (in the flash disk)	1 copy
USB cable	1 pc
Dust cover	1 pc

2 Environmental Requirements of Software

Please read the manual of Windows XP or higher version before reading this section. Windows XP is recommended.

When the operating system is Windows Vista or higher version, please run the software under administrator account. Otherwise the software may not run properly.

Please change the setting in Power Options“Put the computer to sleep” to “Never”. Otherwise it may cause an error while running Prolab.

2.1 PC Requirements

2.1.1 Hardware Requirements


Hardware	Requirements
CPU	Intel 2.5GHz or same level CPU
Memory	No less than 2G
Hard drive	No less than 1G space
USB port	USB2.0
Monitor	Resolution 1024*768 or above Color 16bit or above

2.2 Install Prolab

- Boot your computer.
- Insert the Prolab disk, then open “My computer”.
- Select your CD Rom in the browser.
- Double click “Setup” to install.
- Follow the steps and finish setup, then restart the computer.



Click "Next" to continue.



3 Before Use

3.1 Connect to PC

Please use the USB cable to connect the instrument and PC. Driver installation will automatically start when first time connecting. Please run the software after driver installation is done.

3.2 How To Connect


1. Plug in USB cable

Connect the instrument and PC with USB cable while the PC is on.

2. Turn on the instrument.

Turn on the instrument then run Prolab. The instrument will check if connected.

There will be a dialog pop out as below if not connected or connection error.



“Abort”: Exit the software.

“Retry”: Check again.


“Ignore”: Continue running Prolab. Other functions are available.

3. Initialization

The software will automatically start initializing and adjusting.

4. Interface.



Click  above to create a new measurement. Then select “Photometric”, “Wavelength Scan”, “Time Scan” or “Quantitaion” mode.

5. Shutting down

Close the software first, then turn off the instrument.



If you turn off the instrument first, there will be an error in Prolab. You need to run task manager to end the software.

4 Functions

4.1 Modes

There are 4 modes:

1. Photometric

- (1) Measure the photometric data of the sample.
- (2) Data display as Trans./Abs..
- (3) Photometric data supports up to 26 wavelengths.
- (4) support up to 10 user-defined calculation formula.
- (5) Data printout available.

2. Wavelength Scan

- (1) Get the spectrum of the sample.
- (2) Display as Trans./Abs..
- (3) Repeat scan available.
- (4) Spectrum processing.
- (5) Data printout available.

3. Time Scan


- (1) Get the spectrum of sample varying by time.
- (2) Display as Trans./Abs..
- (3) Repeat scan available.
- (4) Spectrum processing.
- (5) Data printout available.

4. Quantitation

- (1) Supports single-wavelength, dual wavelength and three-wavelength quantitative analysis.
- (2) Supports 1 to 3 times curve fitting.
- (3) Data decimal can be changed.
- (4) Programmable optical gate control.
- (5) Supports data printout.

4.2 Interface

4.2.1 Modules




4.2.2 Modules Intro

1. Menu & Toolbar




Menu & Toolbar




- 1) Provides instrument controls and settings.
- 2) Tool bar are shortcuts for common features.

2. Document Browser

Document Browser shows files saved in Wavelength Scan, Time Scan and Quantitative Analysis mode. Double click to open a file.



- 1) Double click a file to open/reset spectrum.
- 2) Right click a file to open, rename or delete.
- 3) Input keywords to search.



















3. Spectrum Window & Information Window











- 1) Information window shows current data and settings of the instrument.
- 2) In spectrum window, you can use mouse to zoom in and out. Press the left mouse button, drag the mouse from top left to bottom right to draw a square, then release the button. Spectrum in that square will be zoomed in. Drag the mouse the opposite way to zoom out.
- 3) Click "Peaks" to show the peaks in the spectrum.

4. Status Window

Status window shows the current status of the instrument.

4.2.3 Icons


Icon	Function
	New Measurement
	Open Spectrum
	Show/Hide Status
	Show/Hide Spectrum Information
	Wavelength scan / Time scan
	Photometric window
	Quantitation window
	Back to original coordinate
	Auto coordinate
	Y-axis enlarge 2 times
	Y-axis reduce 2 times
	Get/Cancel Axis Data
	Zoom In/Out
	Show/Hide Peaks
	Show/Hide Grid
	Start/Stop

	Set wavelength
	Set 100%
	0%
	Set baseline
	Spectrum Properties
	Print Data
	Trans. / Abs.
	Move sample rack to "Sample" position
	Move sample rack to "Reference" position
	Current status online/offline

5 Software Operation

5.1 Photometric

Photometric work flow






5.1.1 Create Measurement

Create a new measurement.

Select "Files" -> "Create Method" or click  to enter Create Method Window.

1. Measurement summary:

The screenshot shows the 'Create Measurement Method' dialog box with the 'General' tab selected. The 'Measurement' dropdown is set to 'Photometric'. The 'Operator' dropdown is open, showing options: 'Photometric', 'Wavelength Scan', 'Time Scan', and 'Quantitation'. The 'Serial Number' and 'Version Number' fields are empty. The 'Memo' field is a large text area. At the bottom are buttons for 'Default', 'Open', 'Save', 'OK', and 'Cancel'.

- 1) Measure Mode: Choose "Photometric".
- 2) Operator: Input operator's name.
- 3) Serial Number: Shows the serial number of the instrument.
- 4) Version Number: Shows the version of the instrument.
- 5) Memo: Enter a description or notes on measuring conditions.
- 6) Click  to reset.
- 7) Click  to open saved parameters.
- 8) Click  to save the parameters.


2. Instrument Tab:

The screenshot shows the 'Create Measurement Method' dialog box with the 'Instrument' tab selected. The 'Data Mode' dropdown is set to 'Abs.'. The 'Wavelength' field is set to 1500. Below it is a table with columns 'No.' and 'Wl.' containing three rows: A (1000), B (1500), and C (1800). To the right of the table are buttons for 'Add', 'Edit', 'Delete', and 'Clear'. The 'Delay' field is set to 2 s and the 'Integral' field is set to 2 s. The 'Replicate' field is set to 1 and the 'Repeat Mode' dropdown is set to 'Automatic'. The 'Cycle Time' field is set to 1 s. At the bottom are buttons for 'Default', 'Open', 'Save', 'OK', and 'Cancel'.

No.	Wl.
A	1000
B	1500
C	1800

- 1) Data Mode: Data display as Trans. or Abs.
- 2) Wavelength: Input wavelength in , then click to add it in the table.
- 3) Change wavelength: Select the wavelength in the table , then change the value here . Click to finish.
- 4) Delete wavelength: Select the wavelength in the table . Click to delete it.
- 5) Clear: Click to delete all wavelengths.
- 6) Delay: Delay time before measuring. Usually for stabilization.
- 7) Integral: Data integral time.
- 8) Slit: Slit of the instrument. (759S,756S,723S all fixed at 2nm)
- 9) Light source: Shows the switch wavelength of deuterium lamp / tungsten lamp. (Not for Photometric mode./Not for 723S)
- 10) Lamp status: Switch deuterium lamp / tungsten lamp (Not available here).
- 11) Gain: Set a gain to measure sample (Not available here).
- 12) Replicate: Set the number of repeat measurements. The instrument will only scan once when it's 1.
- 13) Cycle time: Set a repetition interval.

3. Custom Formula Tab:



- 1) Title: Input formula title.
- 2) Formula: Use the keyboard below to input formula.

3) Add formula: Input formula here and click to add it in the list.

4) Change formula: Select the formula you want to change

List	
Title	Formula
Formula1	A+B
Formula2	A-B
F3	A*B

,edit it here and click to finish.

5) Delete formula: Select the formula you want to delete

List	
Title	Formula
Formula1	A+B
Formula2	A-B
F3	A*B


, then click to delete it.

6) Clear: Click to delete all formula.

7) Used: Check to calculate after measurement.

8) Display: Display title or formula.

4. Report Tab:





1) Export: Choose "Print Report" or "Save as Microsoft (R) Excel file".


2) General-Date: Add output date.

3) General-Method: Show output method.



4) Other check box not available in Photometric mode.

5.1.2 Start a Measurement

1. Create Photometric measurement as 2.3.1.1.
2. Put the reference sample in the sample cell and click  to set 100% and 0%.
3. Put sample in the sample cell and click  to start measuring.
4. There will be a popout window after the first measurement. Input file name and click OK (or leave it) to save. Data will be saved in this file unless you create a new measurement.




5.1.3 Data Processing

1. Click  to show current method.
2. Click  to print out data.

5.2 Wavelength Scan


Wavelength Scan work flow






5.2.1 Create a Measurement

Select “Files”->“Create Method” or click  to enter Create Method Window.

1. General Tab:



- a) Measure Mode: Choose “Wavelength Scan”.
- b) Operator: Input operator’s name.
- c) Serial Number: Shows the serial number of the instrument.
- d) Version Number: Shows the version of the instrument.
- e) Memo: Enter a description or notes on measuring conditions.
- f) Click  to reset.
- g) Click  to open saved parameters.
- h) Click  to save the parameters.

2. Instrument Tab:


The screenshot shows the 'Create Measurement Method' dialog box with the 'Instrument' tab selected. The dialog has five tabs: General, Instrument, Monitor, Processing, and Report. The 'Instrument' tab contains the following settings:

- Data Mode: Abs. (dropdown menu)
- Wl. Min: 1000 nm (text input)
- Wl. Max: 1800 nm (text input)
- Speed: High (dropdown menu)
- Interval: 1 nm (text input)
- Delay: 0.0 s (text input)
- Replicate: 1 (text input)
- Repeat Mode: Automatic (dropdown menu)
- Cycle Time: 1 s (text input)

At the bottom of the dialog are five buttons: Default, Open, Save, OK, and Cancel.


- 1) Data mode: Data display as Abs., Trans. or Energy.
- 2) WL.Min: Input the start wavelength.
- 3) WL.Max: Input the ending wavelength.
- 4) Speed: Select the scan speed. The faster the more noise appears.
- 5) Interval: Shows data sampling interval according to the scan speed.
- 6) Delay: After pressing the Measure button, measurement is started following the delay time set here.
- 7) Slit: Slit of the instrument.
- 8) Replicate: Set the number of repeat measurements. The instrument will only scan once when it's 1.
- 9) Cycle time: Set a repetition interval.

3. Monitor Tab






- 1) Y Axis: Enter the max and min point of Y axis. The max point should be larger.
- 2) Auto Zoom: Y axis will automatically set by spectrum data.


4. Processing Tab:



- 1) Available: Savitsky-Golaysmooth, Mean smooth, Median smooth and Derivative are available for data processing.


- a) Click a method in the “Available” box then click  to put it in the Selected box.
- b) Click a method in the “Selected” box then click  to remove it.
- 2) Selected: Final data will be calculated with methods in the Selected box. You can set parameters for each method.
- 3) Modification: Click the  to unfold the parameters of each method, click to change it in “Modification”.
- 4) Peak Finding: Automatically find peaks by giving threshold when the scan is complete.

5. Report Tab:





- 1) Output: Print Report or Save as CSV file.
- 2) Output options: Choose the printout data. Check the content in “Properties” button on the left after the scan.
- 3) Add Data: When “Spectrum Data” is checked, you can choose data section to printout. Set the start wavelength, end wavelength and interval in the pop out window, then click OK. Click the “+” to see the data section. Up to 9 sets of data.






- 4) Clear Data: Clear current data section.










5.2.2 Start a wavelength scan











1. Create a new measurement.
2. Put reference sample in the sample cell and click  to adjust baseline.
3. Put sample in the sample cell and click  to start scanning.
4. There will be a popout window after the first measurement. Input file name and click OK (or leave it) to save.



5. Click  to stop the scan.

5.2.3 Data Processing

1. Click  to show the method detail.
2. Click  to print out data.
3. Click  to reset original coordinate.
4. Click  to automatically zoom Y axis.
5. Click  to zoom in Y axis 2 times.
6. Click  to zoom out Y axis 2 times.
7. Click  /  to get/cancel axis data of the cursor.
8. Click  to zoom as set.


9. Click  /  to show/hide peaks.
10. Click  to show/hide grid.
11. Click  to move sample cell.
12. Click  to switch between Trans. and Abs.
13. Click  to run spectrum comparison.
14. Click  to run spectrum calculation.
15. Click  to run Spectrum derivation.
16. Click  to run Spectrum Smoothing.
17. Click  to find peaks.

5.2.4 Model

For users who need to build models for PCA, Prolab will automatically save wavelength data as .DX files in the Prolab directory/DX folder. You can also save wavelength data as .CSV files in Printout data.

5.3 Time Scan


Time scan work flow






5.3.1 Create a new measurement


Select "Files" -> "Create Method" or click  to enter Create Method Window.

1. **General Tab:**




- a) Measure Mode: Choose "Time Scan".
- b) Operator: Input operator's name.
- c) Serial Number: Shows the serial number of the instrument.
- d) Version Number: Shows the version of the instrument.
- e) Memo: Enter a description or notes on measuring conditions.
- f) Click  to reset.
- g) Click  to open saved parameters.
- h) Click  to save the parameters.

2. Instrument Tab:




- 1) Data mode: Data display as Abs., or Trans.
- 2) Wavelength: Input time scan wavelength.
- 3) Unit: Set time unit to sec or ms.
- 4) Time: Set the scan time.
- 5) Interval: Fixed at 0.1 sec.
- 6) Delay: Delay time before scan.
- 7) Slit: Fixed at 8nm.
- 8) Replicate: Set the number of repeat measurements. The instrument will only scan once when it's 1.
- 9) Cycle time: Set a repetition interval.




3. Monitor Tab




- 1) Y Axis: Enter the max and min point of Y axis. The max point should be larger.
- 2) Auto Zoom: Y axis will automatically set by spectrum data.

4. Processing Tab:




1. Available: Savitsky-Golaysmooth, Mean smooth, Median smooth and Derivative are available for data processing.
 - a) Click a method in the “Available” box then click  to put it in the Selected box.
 - b) Click a method in the “Selected” box then click  to remove it.
2. Selected: Final data will be calculated with methods in the Selected box. You can set parameters for each method.
3. Modification: Click the  to unfold the parameters of each method, click to change it in “Modification”.
4. Peak Finding: Automatically find peaks by giving threshold when the scan is complete.

6. Report Tab:







- 1) Output: Print Report or Export to CSV file.
- 1) Output options: Choose the printout data. Check the content in "Properties" button on the left after the scan.
- 2) Add Data: When "Spectrum Data" is checked, you can choose data section to printout. Set the start wavelength, end wavelength and interval in the pop out window, then click OK. Click the "+" to see the data section. Up to 9 sets of data.



- 3) Clear: Clear current segment











5.3.2 Start a Time Scan

1. Create a new time scan measurement as 2.3.3.1.
2. Put reference sample in the sample cell, then click  to set 100%.
3. Click  to set 0%.
4. Put sample in the sample cell, then click  to start scan.
5. There will be a pop-out window after the first measurement. Input file name and click OK (or leave it) to save.





6. Click  to stop the scan.


5.3.3 Data Processing

1. Click  to show detail of current method.
2. Click  to print out data.
3. Click  to reset original coordinate.
4. Click  to automatically zoom Y axis.
5. Click  to zoom in Y axis 2 times.
6. Click  to zoom out Y axis 2 times.
7. Click  to get/cancel axis data of the cursor.
8. Click  to zoom as set.
9. Click  to show/hide peaks.
10. Click  to show/hide grid.


11. Click  to move sample cell.


12. Click  to switch between Trans. and Abs.

13. Click  to do spectrum comparison.

14. Click  to do spectrum calculation.


15. Click  to do spectrum derivation.

16. Click  to do spectrum smoothing.

17. Click  to run peak finding.

5.4 Quantitation


Quantitation work flow






5.4.1 Create a measurement

Select “Files”->“Create Method” or click  to enter Create Method Window.

1. General Tab:



- a) Measure Mode: Choose “Time Scan”.
- b) Operator: Input operator’s name.
- c) Serial Number: Shows the serial number of the instrument.
- d) Version Number: Shows the version of the instrument.
- e) Memo: Enter a description or notes on measuring conditions.
- f) Click  to reset.
- g) Click  to open saved parameters.
- h) Click  to save the parameters.

2. Quantitation Tab:

The screenshot shows the 'Create Measurement Method' dialog box with the 'Quantitation' tab selected. The 'Measurement' section has 'Method' set to 'Wavelength' and 'Wl. Numbers' set to '1'. The 'Unit' is set to '%'. The 'Calibration' section has 'Type' set to 'Conc = f(Abs)', 'Order' set to 'Linear', and both 'Custom Coef' and 'Force Zero' checkboxes are unchecked. The coefficients A0, A1, A2, and A3 are all set to '1'. At the bottom are buttons for 'Default', 'Open', 'Save', 'OK', and 'Cancel'.


- 1) Method: Quantitation method. Only wavelength available now.
- 2) WL Number: Number of wavelengths to analyse with.
- 3) Unit: Concentration unit.
- 4) Type: The type of formula to display.
- 5) Order: Linear, Quadratic and Cubic available.
- 6) Custom Coef: Check the box to customize equation as "Conc = A0 + A1 * X¹ + A2 * X² + A3 * X³".
- 7) Force Zero: Check the box to force the (0,0) point fits the equation.

3. Instrument Tab:

The screenshot shows the 'Create Measurement Method' dialog box with the 'Instrument' tab selected. The 'Data Mode' is set to 'Abs.'. Under 'Ratio', there are three rows: 'No.1: 1 * 1500 nm', 'No.2: 1 * 1500 nm', and 'No.3: 1 * 1500 nm'. The 'Delay' is set to '2 s' and the 'Integral' is set to '2 s'. On the right, 'Replicate' is set to '1', 'Repeat Mode' is set to 'Automatic', and 'Cycle Time' is set to '1 s'. At the bottom are buttons for 'Default', 'Open', 'Save', 'OK', and 'Cancel'.

- 1) Data mode: Choose Abs. or Trans. to display value.
- 2) Wavelength: Input test wavelengths based on WL numbers in Quantitation tab. When WL number is 3, the number of 3 wavelengths need to be increasing or decreasing.
- 3) Delay: Delay time before scan.
- 4) Integral: Data integral time.
- 5) Slit: Fixed at 8nm.
- 6) Replicate: Set the number of repeat measurements. The instrument will only scan once when it's 1.
- 7) Cycle time: Set a repetition interval.

4. Report Tab:


















- 1) Export: Print Report or Export to file.
- 2) Date: Export with date.
- 3) Spectrum: Export with spectrum.
- 4) Method: Export with method detail.
- 5) Calibration: Export with equation.
- 6) Standard Data: Export with standard data.
- 7) Sample Data: Export with sample data.
- 8) Others are not available in Quantitation.


5.4.2 Build Calibration Curve

1. In the Standards window, you can modify sample name, description and concentration; add or delete sample; check the value of samples.

No.	Name	Memo	Conc.	Abs.	Use	Cell
1			1	0.145	<input checked="" type="checkbox"/>	-
2			2	0.269	<input checked="" type="checkbox"/>	-
3			3	0.34	<input checked="" type="checkbox"/>	-
4			4	0.466	<input checked="" type="checkbox"/>	-

Measure Edit Insert Delete Build

- 1) To change sample name, description and concentration: Click , then double click in the table to modify the content you want. Then click  to confirm.
 - 2) Measure a standard sample: Click to select a sample in the table, then click  button. The instrument will start measurement.
 - 3) Add Sample: Click , then click . There will be another line in the standards window. Click  to finish.
 - 4) Delete Sample: Click  and click a line you want to delete, then click . Click  to finish.
 - 5) Choose the sample data needed in curve calculation: Click , then click the check mark in  row if you want to use this data for calculation.
 - 6) Set sample cell: Click  and set sample cell of samples in  column. "-" means do not move sample cell. Click  to finish.
2. Click  to build the curve of standard sample when finishing measurement.









5.4.3 Measuring unknown samples




When the regression curve is created, you can start measuring the sample. Operate the test sample in samples window as below. There are functions in the sample window: Measure, Modify, Delete and Clear.

Sample Data				
No.	Name	Memo	Abs.	Conc.
1			0.0005	0.01
2			0.0005	0.01
3			0.0005	0.00
4				
5				
6				
7				
8				
9				
10				
11				



Measure Edit Delete Clear

- 1) Change sample name & note: Click  button, then double click the frame you want to modify. Click  to confirm the modification and back to test sample window.
- 2) Measure sample value: Click a sample value frame, then click  button to measure the sample. The value of the sample will be in "Abs." and "Conc." column.
- 3) Delete sample: Click  button, then click the line you want to

delete and click  button to delete the sample. Click  to confirm the modification and back to test sample window.


- 4) Clear sample list: Click  button, then click  button. Click  to confirm the modification and back to test sample window.


5.4.4 Data Processing

1. Click  to show detail of current measurement.
2. Click  to print out data.

5.5 General Operation

5.5.1 Wavelength




When the instrument is not measuring (shows **Ready** below), click  to open wavelength dialog. Input the wavelength you want then click **OK** to go.


Wavelength range:

- S430: 1000nm~1800nm


5.5.2 Set 100%

When the instrument is not measuring (shows **Ready** below), click  to set 100%. Status shows **Adjust Full Data** **Set 100** below. When it shows **Ready**, means it is done.


5.5.3 Set 0%

When the instrument is not measuring (shows **Ready** below), click  to set 100%. Status shows **Adjust Zero** **Set Zero** below. When it shows **Ready**, means it is done.


5.5.4 Move sample rack

When the instrument is not measuring (shows **Ready** below), click  to move sample rack. Status shows **Shifting Sample Cell** below. When it shows **Ready**, means it is done.



5.5.5 Switch T/A

When the instrument is not measuring (shows **Ready** below), click  to switch T/A.

Switch current value




- Switch spectrum data




5.5.6 Locate file



To locate test files, you can input keyword in keyword blank.





5.5.7 Data management

Click File-Data management to enter Data management. In data management you can quickly delete, export, hide in batch.




1. Refresh: Select type, status and input keyword, then click  to show the fitting files.
2. Batch Export: Select the files you need, then click  to export CSV file. This will export all data.

3. Batch Print: Select the files you need, then click  to print one by one. It will print all data in the file.
4. Change status: Switch between Display/Hide. Hidden files won't be seen in the list outside Data management.
5. Delete: Select the files you want to delete, then click  to delete those files.


5.5.8 Characteristic Peak


Click Data Processing – Characteristic Peak to enter Characteristic Peak dialog.



1. WL List: The wavelengths and sample cell position.
2. WL Para.: Modify wavelength and sample cell position here.
3. Data mode: Display data in Abs. or Trans.
4. Range: Choose a scan range around testing wavelength.
5. Baseline: Do baseline scan around every testing wavelength according to Range value. Scan sample cell position 4 as default:
6. Measure: Start scanning peaks.

5.5.9 Arithmetic

Arithmetic is to do addition, subtraction, multiplication and division operations of the same type of spectrum. Click  to enter Arithmetic.




1. Functions: Select arithmetic, calculate, save and export.


Button	Function
	Select addition, subtraction, multiplication or division
	Click to calculate the two spectrums.
	Click to save a BMP file.
	Click to print.

1. Spectrum 1 & Files 1: Shows the select spectrum in files 1.
2. Spectrum 2 & Files 2: Shows the select spectrum in files 2.
3. Spectrum type: Click to browse the same type of spectrum.
4. Results: Show the last result.
Result spectrum = 【Target spectrum 1】 +/-/*/* 【Target spectrum 2】.

5.5.10 Spectrum Compare


This function is to compare the same type of spectrum. Click  to enter Compare.

1. Interface:




- a) Type: Choose the spectrum type.
- b) Files: Shows the certain type of spectrums. Hold Ctrl to select multiple spectrums.


2. Click OK to compare selected spectrums.





- a) Using different colors to distinguish different spectrums in spectrum window.
- b) You can check or uncheck a spectrum in the selected box.
- c) Functions: Save and Print spectrum compare.

5.5.11 Find Peak




This function is to quickly find peaks of spectrum. Click  to set threshold according to current spectrum. Higher peaks need larger Y axis threshold; wider peaks need larger X axis threshold.

Click  to show peaks. The icon turns to . Click it to hide peaks. Spectrum information below will show peak info.


5.5.12 Smooth




Smooth is to reduce the noise of spectrum. Click  to set smoothing parameters. Select type, Smoothing order, Number of points and Number of times then click "OK" to see the effect.

Type	Smoothing order	Number of points	Number of times
Savitzky-Golay	The highest power of the polynomial	Set the number of points to be used in calculation.(odd number)	Set the number of smoothing operations.
Mean	-----	Set the number of points to be used in calculation.	Set the number of smoothing operations.
Median	-----	Set the number of points to be used in calculation.	Set the number of smoothing operations.

5.5.13 Derivative




Derivative operation on spectrums is to enhance the resolution of peaks. Derivation can distinguish various disturbances affecting the shape of the spectrum peaks. Usually combining the smoothing operation.


Click  to open Derivative parameters window. Set Derivative order and click "OK" to see the result.

5.5.14 Instrument Parameter

This is to change the save path, file name and spectrum type. Click “Settings”


→“Options”  to open option window.

1. General tab




- 1) Default Method: Instrument use default parameters.
- 2) New method: Instrument uses specified parameters.

2. Save Tab:




- 1) Automatic Increase after Scan: Automatically add a number suffix to file names.
- 2) Start: Set the start number.
- 3) Digit: Set number digits.
- 4) Reset after rename: Auto reset number when using another name.


3. Spectrum Tab:



- 1) Curve color: Sets curve color.
- 2) Serie width: Sets curve width.
- 3) Point color: Sets dot color.
- 4) Point width: Sets dot width.
- 5) Point height: Sets dot height.
- 6) Point style: Sets dot shape as rectangle, circle, triangle, down triangle, cross, diacross, star and diamond.




- 1) Grid color: Sets the table color in "Spectrum information" window.
 - 2) Grid width: Sets the curve width in "Spectrum information" window.
 - 3) Grid style: Sets table style as solid, dash, dot, dashdot, dashdotdot and clear in "Spectrum" tab.
 - 4) Panel color: Sets the background color in "Spectrum information" window.
 - 5) Wall color: Sets the coordinate board color.
 - 6) Wall Transparent: Check it to make the board transparent.
4. Auto Reference Tab:



When this item is checked, the reference data is corrected before each measurement in "wavelength scan".

5.5.15 Rename & Delete Files


Right click on a file in file browser. You can delete or rename in the pop-out menu.



6 Appendix


6.1 Quantitative analysis wavelength method

6.1.1 Single Wavelength



Abs. A_1 is the value on the curve at λ_1 .


6.1.2 Double Wavelengths



A_1 and A_2 are the Fluorescence at λ_1 and λ_2 .

$$A = A_2 - A_1$$

6.1.3 Triple Wavelengths



A_1 , A_2 and A_3 are the Fluorescence at λ_1 , λ_2 and λ_3 .

$$= \frac{(A_1 - A_2) \times \lambda_3 + (A_2 - A_3) \times \lambda_1}{\lambda_1 - \lambda_3}$$

(It has to be $\lambda_1 > \lambda_2 > \lambda_3$ or $\lambda_1 < \lambda_2 < \lambda_3$)

6.2 DETAILS ON QUANTITATIVE

Prolab provides 3 calibration types: Linear working curve, Quadratic working curve and Cubic working curve. All of them are not forced through the 0 coordinates.

6.2.1 Linear Working Curve (1st order)

The calculation formula is as follow:

$$A = C \times K_1 + K_0$$

Where,

C : Concentration of each sample (input value)

A : Abs. of each sample (measured value)

K_1 and K_0 are calculated by the least squares method.

Suppose there are n data points (A_n , C_n), then

$$K_1 = \frac{\sum_{i=1}^n A_i C_i - \frac{1}{n} \sum_{i=1}^n A_i \cdot \sum_{i=1}^n C_i}{\sum_{i=1}^n A_i^2 - \frac{1}{n} (\sum_{i=1}^n A_i)^2}$$

$$K_0 = \frac{\sum_{i=1}^n C_i}{n} - K_1 \times \frac{\sum_{i=1}^n A_i}{n}$$

6.2.2 Quadratic Working Curve (2nd order)

The calculation formula is as follow:

$$= a_2 \times C^2 + a_1 \times C + a_0$$

Where,

C: Concentration of standard sample

A : Abs. of each sample (measured value)

K_n are calculated by the least squares method

Suppose there are n data points(A_n , C_n), then:

$$K_2 = \frac{S(A^2C)S(AA) - S(AC)S(AA^2)}{S(AA)S(A^2A^2) - [S(AA^2)]^2} \quad (\text{Formula F5-5})$$

$$K_1 = \frac{S(AC)S(A^2A^2) - S(A^2C)S(AA^2)}{S(AA)S(A^2A^2) - [S(AA^2)]^2} \quad (\text{Formula F5-5})$$

$$K_0 = \frac{\sum_{i=1}^n C_i}{n} - K_1 \frac{\sum_{i=1}^n A_i}{n} - K_2 \frac{\sum_{i=1}^n A_i^2}{n} \quad (\text{Formula F5-7})$$

6.2.3 The correlation coefficient


The correlation coefficient R represents how the regression curve fitting.

Suppose there are n data points(,):

$$= \frac{\sum_{i=1}^n A_i C_i - \frac{\sum_{i=1}^n A_i \cdot \sum_{i=1}^n C_i}{n}}{\sqrt{\left(\sum_{i=1}^n A_i^2 - \frac{(\sum_{i=1}^n A_i)^2}{n} \right) \left(\sum_{i=1}^n C_i^2 - \frac{(\sum_{i=1}^n C_i)^2}{n} \right)}} \quad (\text{Formula F5-13})$$

6.3 Derivative Operation on Spectrum

The derivative of a function of a real variable measures the sensitivity to change of a quantity (a function or dependent variable) which is determined by another quantity (the independent variable).



Derivative of the function

There are many ways of derivative operation on spectrum. Since the x-axis(time axis or the wavelength axis, etc.) of the original spectral data are equally spaced, then

First order derivative:

$$\frac{dy}{dx} = \frac{y_{i+1} - y_i}{\Delta x}$$

Second order derivative:

$$\frac{d^2y}{dx^2} = \frac{y_{i+1} - 2y_i + y_{i-1}}{\Delta x^2} \quad (\text{Formula F7-2})$$

Where:

y: photometric value

x: wavelength, time, etc.

Derivative spectra not only can eliminate baseline drift or flat background interference, but also can provide a higher resolution than the original spectrum.




Fig. 1




Fig. 2




Fig. 3

In Fig.1, there is a clear alternation of peaks. In Fig.2, the acromion is higher after derivative. Second order derivative spectrum is clearer. In Fig.3, the original spectrum two curves are seriously overlapping, but in $n=2/4$ the peaks are clearer.

Higher order derivative can eliminate the low order background curves. The spectrum shape is complicated after derivative, but it raises the resolution and sensitivity.

6.4 Smoothing

The basic idea of smoothing is to map a smooth point, then depict a number of points around the smooth point to be "fit" or "average" or "sort" in order to obtain the best estimate of the value of the smooth point to eliminate random noise. With modern analytical instruments increasing speed and automation, multiple accumulate and smoothing technology has become a common method of noise reduction.

Prolab provide 3 smoothing methods: Savitsky-Golay, Mean and Median.

6.4.1 Savitzky–Golay

A **Savitzky–Golay filter** is a digital filter that can be applied to a set of digital data points for the purpose of smoothing the data, that is, to increase the signal-to-noise ratio without greatly distorting the signal. This is achieved, in a process known as convolution, by fitting successive sub-sets of adjacent data points with a low-degree polynomial by the method of linear least squares. When the data points are equally spaced an analytical solution to the least-squares equations can be found, in the form of a single set of "convolution coefficients" that can be applied to all data sub-sets, to give estimates of the smoothed signal, (or derivatives of the smoothed signal) at the central point of each sub-set.

6.4.2 Mean smoothing

Median smoothing is to sort the selected data (the number of data points is odd), then take the intermediate value as the smoothed value.

6.4.3 Median smoothing

Median smoothing is to sort the selected data (the number of data points is odd), then take the intermediate value as the smoothed value.